



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***Eastex Environmental Lab, Inc.***  
*35 Eastex Lane, Coldspring TX 77331 United States*

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025: 2017**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Chemical Testing***  
*(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

*Initial Accreditation Date:*

August 7, 2020

*Issue Date:*

August 7, 2020

*Expiration Date:*

November 30, 2022

Tracy Szerszen  
President

*Accreditation No.:*

78075

*Certificate No.:*

L20-466

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjllabs.com](http://www.pjllabs.com)*



# Certificate of Accreditation: Supplement

## Eastex Environmental Lab, Inc.

35 Eastex Lane Coldspring TX 77331 United States  
 Contact Name: Tiffany Guerrero Phone: 936-653-3249

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Chemical <sup>F</sup>	Cannabis Plant Material, Concentrates, Infused Products	<b>Cannabinoids:</b> CBDV CBDA CBGA CBG CBD THC THC delta 8 CBC THCA CBD CBN CBCA	HPLC UV	D.L. = 0.001 8 % D.L. = 0.000 4 % D.L. = 0.000 6 % D.L. = 0.000 4 % D.L. = 0.000 4 % D.L. = 0.000 6 % D.L. = 0.000 8 % D.L. = 0.000 4 % D.L. = 0.000 5 % D.L. = 0.000 4 % D.L. = 0.000 4 % D.L. = 0.001 4 %
		<b>Cannabinoids:</b> CBDV CBDA CBGA CBG CBD THC THC delta 8 CBC THCA CBD CBN CBCA	HPLC UV	D.L. = 0.018 mg/g D.L. = 0.004 mg/g D.L. = 0.006 mg/g D.L. = 0.004 mg/g D.L. = 0.004 mg/g D.L. = 0.006 mg/g D.L. = 0.008 mg/g D.L. = 0.004 mg/g D.L. = 0.005 mg/g D.L. = 0.004 mg/g D.L. = 0.004 mg/g D.L. = 0.014 mg/g
		<b>Residual Solvents:</b> 1,2-Dichloroethane Benzene Chloroform Dichloromethane Trichloroethylene Acetone Acetonitrile Ethanol Ethyl Acetate Ethyl ether Heptane Hexane Isopropyl alcohol Methanol Pentane Toluene Total Xylenes (ortho-, meta-, para-)	GC/MS	D.L. = 0.44 µg/g D.L. = 0.43 µg/g D.L. = 0.45 µg/g D.L. = 0.56 µg/g D.L. = 0.49 µg/g D.L. = 0.53 µg/g D.L. = 0.81 µg/g D.L. = 0.71 µg/g D.L. = 0.55 µg/g D.L. = 0.43 µg/g D.L. = 0.49 µg/g D.L. = 0.57 µg/g D.L. = 0.71 µg/g D.L. = 0.71 µg/g D.L. = 0.92 µg/g D.L. = 0.51 µg/g D.L. = 1.01 µg/g



# Certificate of Accreditation: Supplement

**Eastex Environmental Lab, Inc.**  
35 Eastex Lane Coldspring TX 77331 United States  
Contact Name: Tiffany Guerrero Phone: 936-653-3249

*Accreditation is granted to the facility to perform the following testing:*

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Chemical <sup>F</sup>	Cannabis Plant Material	<b>Water Testing:</b> Loss on Drying (Moisture Content)	Gravimetric	0.1 % to 99 %

1. The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this testing at its fixed location.

